

Statement of Product Conformity

**BM Polyco Ltd
Crown Road
Enfield
EN1 1TX
Great Britain**

Certified to ISO 9001:2008

declare that the new Medical Device and PPE described hereafter



FINITE P INDIGO AF

**Non-sterile, powder-free, synthetic nitrile rubber
disposable examination glove**

is in conformity with the provisions of Council Directive 93/42/EEC and with National Standards transposing harmonised standards EN455 part 1:2000, EN455 part 2:2000 and EN455 part 3:2000 and is self-certified as a Class 1 non-sterile medical device

is in conformity with the provisions of Council Directive 89/686/EEC and where such is the case, with National Standards transposing harmonised standards EN374-1:2003, EN374-2:2003, EN374-3:2003 and EN420:2003.

are identical to the PPE which is the subject of EC certificate of conformity No 3766 issued by SATRA Technology Centre, Wyndham Way, Telford Way, Kettering, Northamptonshire, NN16 8SD, United Kingdom (Notified Body number 0321).

Conform to the provisions of Council Directive 89/686/EEC Article 11B under the supervision of the Notified Body, LRQA Ltd, Hiramford, Middlemarch Office Village, Siskin Drive, Coventry, CV3 4FJ (Notified Body number 0088).

**MICRO ORGANISM
HAZARDS
EN374-2**



**CHEMICAL
HAZARDS
EN374-3**



CE 0088

Allergens: does not contain chemical accelerators
does not contain natural rubber latex protein

Glove Care

Glove Care: To get best service from your gloves ensure that all contaminant is removed

Storage: Store in dry conditions away from direct sunlight and heat.
You are advised to retain this packaging for reference

Range Available: sizes and codes

Sizes	6.5	7.5	8.5	9.5
Code	MFNP100/01	MFNP100/02	MFNP100/03	MFNP100/04