# MD 100 Photometer

# Precise Water Analysis in High-Quality Design

# Small I Mobile I Rapid

The MD 100 uses high quality interference filters with long-life LEDs as a light source without any moving parts in a transparent sample chamber.

The units supply accurate, reproducible results very quickly. Other major advantages include ease of operation, ergonomic design, compact dimensions and safe handling.

The calibration and software-based adjustment options mean that the MD 100 is also suitable for use as a testing instrument.

The tests are conducted using either Lovibond® tablet reagents with long-term stability and a guaranteed minimum 5 or 10 year shelf life, VARIO powder reagents or using liquid reagents.

Please see pages 78 onwards for reagents (order codes)

# Highlights

- Scroll memory
- Automatic switch-off
- Real-Time-Clock and date
- · Calibration mode
- Backlit display
- Storage function
- One Time Zero (OTZ)
- Waterproof\*)

\*) as defined in IP 68, 1 hour at 0.1 meter



Single-Parameter		Single-Parameter		4in1	
Test	Code	Test	Code	Test Code	
<b>Aluminium,</b> tablet reagents 0.01 - 0.3 mg/l Al	27 62 00	<b>Phosphate</b> , tablet reagents 0.05 - 4.0 mg/l PO <sub>4</sub>	27 60 40	<b>Chlorine, pH, Cyanuric acid,</b> 27 80 70 <b>Alkalinity-M</b> , tablet reagents <b>(OTZ)</b>	
<b>Aluminium</b> , powder reagents 0.01 - 0.25 mg/l Al	27 62 05	<b>Phosphate</b> , powder reagents 0.06 - 2.5 mg/l PO <sub>4</sub>	27 60 45	0.02 - 6.0 mg/l Cl₂ / 0,1 - 10 mg/l Cl₂* 6.5 - 8.4 pH ; 0 - 160 mg/l cyanuric acid 5 - 200 mg/l CaCO₃ (TA)	
Ammonia, tablet reagents 0.02 - 1.0 mg/l N	27 60 60	<b>Silica</b> , tablet reagents $0.05 - 4.0 \text{ mg/l SiO}_2$	27 61 10	Chlorine, pH, Cyanuric acid, 27 80 7 Alkalinity-M (total)	
Ammonium, powder reagents D.01 - 0.8 mg/l N	27 60 65	<b>Silica LR</b> , powder reagents 0.1 - 1.6 mg/l SiO₂	27 61 15	liquid reagent for chlorine and pH <b>(OTZ)</b> 0.02 - 4 mg/l Cl₂ / 6.5 - 8.4 pH 0 - 160 mg/l cyanuric acid / 5 - 200 mg/l CaCO₃ (T	
<b>Chloride</b> , tablet reagents ∂.5 - 25 mg/l Cl¯ 5 - 250 mg/l Cl¯ (by dilution)	27 61 80	Silica HR, powder reagents 1 - 90 mg/l SiO <sub>2</sub>	27 61 16		
Chlorine, tablet reagents (OTZ) 0.01 - 6.0 mg/l Cl <sub>2</sub> / 0.1 - 10 mg/l Cl <sub>2</sub>	27 60 00	Suspended solids no reagents required 0 - 750 mg/l TSS	27 61 50	5in1	
Chlorine, liquid reagent (OTZ) 0.02 - 4 mg/l Cl <sub>2</sub>	27 60 05	<b>Urea</b> , tablet reagents 0.1 - 2.5 mg/l Urea	27 62 10	Chlorine, pH, Cyanuric acid, 27 80 80	
Chlorine DUO, for 2 types of reagen 1) Tablet reagents 0.01 - 6.0 mg/l Cl₂ / 0,1 - 10 mg/l C 2) Powder reagents 0.02 - 2.0 mg/l Cl₂ (ø 24 mm glass 0.1 - 8.0 mg/l Cl₂ (ø 10 mm multi	27 60 20 Cl <sub>2</sub> * 27 60 25 vial) <b>vial-2</b> )	0.2 - 5 mg/l Urea (by dilution)		Alkalinity-M, Calcium hardness tablet reagents (OTZ) 0.02 - 6.0 mg/l Cl <sub>2</sub> / 0,1 - 10 mg/l Cl <sub>2</sub> * 6.5 - 8.4 pH; 0 - 160 mg/l cyanuric acid 5 - 200 mg/l CaCO <sub>3</sub> (TA); 0 - 500 mg/l CaCO <sub>3</sub> (Cal	
<b>Chlorine</b> , powder reagents 0.02 - 2.0 mg/l Cl <sub>2</sub> (ø 24 mm glass via 0.1 - 8.0 mg/l Cl <sub>2</sub> (ø 10 mm <b>multi via</b>		2in1		6in1	
Chlorine HR (Potassium iodide) tablet reagents 5 - 200 mg/l Cl $_2$ (ø 16 mm round vial	27 61 70 & adapter)	Chlorine, pH, tablet reagents (OTZ)	27 80 20	Chlorine, Bromine, pH, 27 80 90	
Chlorine dioxide, tablet reagents 0.02 - 11 mg/l ClO <sub>2</sub>	27 60 30	0.01 - 6.0 mg/l Cl <sub>2</sub> / 0,1 - 10 mg/l Cl <sub>2</sub> * 6.5 - 8.4 pH	27 80 25	Cyanuric acid, Alkalinity-M, Calcium hardness, tablet reagents (OTZ) 0.02 - 6.0 mg/l Cl <sub>2</sub> / 0,1 - 10 mg/l Cl <sub>2</sub> *	
Chlorine dioxide, powder reagents 0.04 - 3.8 mg/l ClO <sub>2</sub>	27 60 35	<b>Chlorine, pH</b> , liquid reagent <b>(OTZ)</b> 0.02 - 4 mg/l Cl <sub>2</sub> / 6.5 - 8.4 pH	27 80 25	0.05 - 13 mg/l Br ; 6.5 - 8.4 pH 0 - 160 mg/l cyanuric acid ; 5 - 200 mg/l CaCO₃ (T/ 0 - 500 mg/l CaCO₃ (CaH)	
<b>COD</b> , tube tests, without reagents 0 - 150 mg/l O <sub>2</sub> (ø 16 mm) 0 - 1500 mg/l O <sub>2</sub> (ø 16 mm) 0 - 15000 mg/l O <sub>2</sub> (ø 16 mm)	27 61 20	<b>Chlorine, pH</b> , powder reagents for chlorine $0.02 - 2.0 \text{ mg/l Cl}_2$ (Ø 24 mm glass vial $0.1 - 8.0 \text{ mg/l Cl}_2$ (Ø 10 mm <b>multi via</b> l $6.5 - 8.4 \text{ pH}$	)		
<b>Copper</b> , tablet reagents 0.05 - 5.0 mg/l Cu	27 60 80	'			
<b>Copper</b> , powder reagents 0.05 - 5.0 mg/l Cu	27 60 85	3in1		MD 100 Boiler Water	
Hardness, total, tablet reagents 2 - 50 mg/l CaCO <sub>3</sub> 20 - 500 mg/l CaCO <sub>3</sub> (by dilution)	27 61 90	Chlorine, pH, Cyanuric acid	27 80 10	Aluminium, Chloride, Copper, 27 62 30	
Hazen, no reagents required 0 - 500 mg/l Pt-Co	27 61 60	tablet reagents <b>(OTZ)</b> 0.01 - 6.0 mg/l Cl <sub>2</sub> / 0,1 - 10 mg/l Cl <sub>2</sub> * 6.5 - 8.4 pH; 0 - 160 mg/l cyanuric ac		DEHA, Hydrazine, Iron, Oxygen (dissolved), Phosphate,	
<b>lron</b> , tablet reagents 0.02 - 1.0 mg/l Fe	27 60 50	Chlorine, pH, Cyanuric acid liquid reagent for chlorine and pH (OT	27 80 15	<b>Polyacrylate, Silica</b> (delivery without reagents)	
Iron TPTZ, powder reagents 0.02 - 1.8 mg/I Fe	27 60 55	0.02 - 4 mg/l Cl <sub>2</sub> / 6.5 - 8.4 pH 0 - 160 mg/l cyanuric acid	<b>2</b> )	AAD 100 Cooling Weber	
lron, powder reagents 0.02 - 3.0 mg/l Fe	27 60 56	Chlorine, pH, Alkalinity-M tablet reagents (OTZ)	27 80 60	MD 100 Cooling Water	
Fluoride, without reagents 0.05 - 2.0 mg/l F	27 60 90	0.01 - 6.0 mg/l Cl₂ / 0,1 - 10 mg/l Cl₂* 6.5 - 8.4 pH ; 5 - 200 mg/l CaCO₃ (TA)		Aluminium, Bromine, Chlorine, 27 62 40 Chlorine HR, Chlorine dioxide, Copper, Iron, Iron in Mo, Molybdate LR,	
Manganese LR, tablet reagents 0.2 - 4.0 mg/l Mn Manganese LR, powder reagents	27 61 00 27 61 05	Chlorine, pH, Alkalinity-M (total) liquid reagent for chlorine and pH (OT	27 80 65 <b>'Z)</b>	Molybdate HR, Ozone, Polyacrylate, Sulphate, Triazoles, Zinc (delivery without reagents)	
0.01 - 0.7 mg/l Mn  Manganese HR, powder reagents	27 61 05	0.02 - 4 mg/l Cl <sub>2</sub> / 6.5 - 8.4 pH 5 - 200 mg/l CaCO <sub>3</sub> (TA)	27.00.00		
0.1 - 18 mg/l Mn  Molybdenum LR	27 61 40	Chlorine LR, Chlorine HR, Chlorine dioxide*, tablet reagents 0.01 - 6.0 mg/l Cl <sub>2</sub>	27 80 00	* Delivery without reagents for measuring range 0.1 - 10 mg/l $\text{Cl}_2$	
Powder reagents / reagent solution 0.03 - 3.0 mg/l Mo (mixing cylinder required, not included)		5 - 200 mg/l Cl <sub>2</sub> (ø 16 mm round vial) 0.02 - 11 mg/l ClO <sub>2</sub>		# Where chlorine and chlorine dioxide are present together, they may be determined quantitatively as a single figure.	
<b>Molybdenum HR</b> , powder reagents 0.3 - 40 mg/l Mo	27 61 41				
Molybdenum, tablet reagents	27 61 42				

# MD 100 Photometer



# Scroll Memory (SM)

To avoid unnecessary scrolling for the required test method, the instrument memorizes the last method used before switching off the instrument. When the instrument is switched on again, the scroll list comes up with the last used test method first.

# **Delivery Content**

- Instrument in carrying case
- 4 micro batteries (AAA)
- 3 Round vials (glass) with lid
- 1 stirring rod & 1 brush
- Tablet reagents and/or liquid reagents or VARIO Powder reagent
- Warranty information
- Certificate (COC)
- Instruction Manual

## Zero Setting (OTZ)

For certain versions of the instrument it is not necessary to zero the instrument each time. The zero setting is held in memory until the device is turned off. (One Time Zero - OTZ). The zero setting can be confirmed whenever it is required.

#### Manufacturers Test Certificate M

Besides the "Certificate of Compliance" which is supplied with the MD 100, manufacturers test certificates M are available at cost on request. Manufacturers test certificates M are individually supplied per instrument and per method.

The manufacturers test certificate M has to be ordered together with the new instrument and cannot be delivered at a later stage.

### N.I.S.T Traceability

The instrument has a factory calibration, which is related to international standards which are not N.I.S.T traceable. The instrument may be calibrated by the user in a "user calibration mode" with N.I.S.T traceable standards.

(N.I.S.T. = National Institute of Standards and Technology)

## **Technical Data**

	er de	
Optics	LEDs, interference filters (IF) and photo sensor in transparent sample chamber. Depending on the version, up to 3 different interference filters are used. Wavelength specifications of interference filters: 430 nm $\Delta \lambda = 5$ nm 530 nm $\Delta \lambda = 5$ nm 560 nm $\Delta \lambda = 5$ nm 580 nm $\Delta \lambda = 5$ nm 610 nm $\Delta \lambda = 6$ nm 660 nm $\Delta \lambda = 6$ nm 660 nm $\Delta \lambda = 5$ nm	
Wavelength Accuracy	± 1 nm	
Photometric Accuracy <sup>4)</sup>	3 % FS (T = 20 °C – 25 °C)	
Photometric Resolution	0.01 A	
Power Supply	4 micro batteries (AAA), capacity approx. 17 hours or 5000 tests	
Auto - OFF	automatic switch-off	
Display	backlit LCD (on keypress)	
Storage	internal ring memory for 16 data sets	
Interfaces	infrared interface for test data transfer	
Additional feature	real time clock and date	
Calibration	factory calibration and user calibration. Reset to factory calibration possible	
Dimensions	155 x 75 x 35 mm (L x W x H)	
Weight	basic unit approx. 260 g	
Environmental conditions	temperature: 5–40°C rel. humidity: 30–90 % (non condensing)	

## **CE-Conformity**

<sup>4)</sup> tested with standard solutions



#### Accessories Code Item Set of 12 round vials with lid 19 76 20 Height 48 mm, Ø 24 mm Set of 5 round vials with lid 19 76 29 Height 48 mm, Ø 24 mm Set of 10 round vials with lid 19 76 65 Height 90 mm, Ø 16 mm Adapter for round vials ø 16 mm 19 80 21 90 Set of 12 plastic vials (PC), with lid 19 76 00 **"Multi"-Type 2**, Ø 10 mm Vial stand for 6 round vials 41 89 51 Ø 24 mm, acrylic glass Vial stand for 10 vials 41 89 57 (Ø 16 mm or □ 13,5 mm), acrylic glass Mixing cylinder, 25 ml, with stopper 19 80 26 50 required accessory for molybdenum LR test with MD 100 (276140) Membrane filter set for use when 36 61 50 preparing samples, 25 membrane filters, 0,45 µm, 2 syringes 20 ml

Cleaning cloth for vials	19 76 35
Set of 12 sealing rings for round vial ø 24 mm	19 76 26
4 micro batteries (AAA)	19 50 026
Measuring beaker, volume 100 ml	38 48 01
Plastic funnel with handle	47 10 07
Plastic stirring rod, 13 cm length	36 41 00
Plastic stirring rod, 13 cm length, (10 pc.)	36 41 20
Plastic stirring rod, 10 cm length	36 41 09
Plastic stirring rod, 10 cm length, (10 pc.)	36 41 30
Infrared data transfer module IRiM	21 40 50



Please see pages 78 onwards for reagents (order codes)





#### Data transfer

The optional available IRiM (infrared interface module) uses modern infrared technology to transmit measurement data from the MD 100 photometer to one of 3 optional interfaces. These interfaces can be used to connect to a PC, a USB printer<sup>1)</sup> or alternatively a serial printer<sup>2)</sup>.

The unit is supplied complete with data logging software providing easy and rapid transfer of data to the PC. As an option, the data can be saved as an Excel sheet or a .txt file.

Measurement data can quickly be printed out, using a specified<sup>1)</sup> USB or alternatively a printer with a serial plug-in connected to the IRiM.

Applicable for the following operating systems: Windows® XP, Windows® Vista and Windows® 7/10.

<sup>1)</sup> USB printer: HP Deskjet 6940 ; <sup>2)</sup> each ASC**II** printer Windows<sup>®</sup> is a registered Trademark of Microsoft Corporation

## Verification Standard Kit

The verification standard kit for the MD 100 is designed to assure the user of the accuracy and the reliability of the results related to the integrated wave lengths.

The kit contains one zero standard, 6 different vials for checking 6 different wave lengths and allows checking the complete range of MD 100 photometers.

The shelf life of the verification standard kit is two years from the date of production, provided that storage and use are in accordance with the instructions provided.

21 56 70

Verification Standard Kit

### Reference Standard Kit for MD 100

The reference standards are designed to check the accuracy and the reliability of the results.

It is not possible to calibrate the photometer with the reference standards.

The shelf life of reference standards is two years from the date of production, provided that storage and use are in accordance with the instructions provided.

provided.	
<b>Kit Chlorine</b> for instruments with tablet / liquid reagent	27 56 50
0.2* and 1.0* mg/l	
Kit Chlorine for instruments	27 56 55
with tablet / liquid reagent	
0.5* and 2.0* mg/l	
Kit Chlorine for instruments	27 56 56
with tablet / liquid reagent	
1.0* and 4.0* mg/l	
Kit Chlorine for instruments	27 56 60
with powder reagent (VARIO)	
0.2* and 1.0* mg/l	
Kit pH for instruments	27 56 70
with tablet / liquid reagent	
7.45* pH	

\* Approximate figure, actual figure specified in Certificate of Analysis

