



Certificate of Analysis

Electrode Care and Maintenance

Electrode Storage Solution

Product No: 1145514
Lot No: ESS15H1
Expiry date: 28/08/2017

Assay: 2.99 M KCl
Date of measurement: 18/08/2015

Specification:
2.90 - 3.10M KCl

pH Specification:
3.9 - 4.1 @20°C

Method:
Tested by argentometric titration , to in-house method TPATPPT1

Traceability:
This solution was checked by means of Silver Nitrate. This volumetric standard is directly traceable to a Standard Reference Material of National Institute of Standards and Technology (USA), SRM 919b Sodium Chloride.

Products are manufactured under an NSAI registered I.S EN ISO9001:2008 Quality System, registration no: 19.2769

Date of issue of the certificate: 18/08/2015

Q.C. Supervisor Margaret Dornan

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-15F3

TRUEscience by Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com