

neodisher[®] LaboClean FT

Liquid, alkaline cleaning agent
for use in special washing machines



Main fields of application: Automated cleaning of laboratory glassware in bacteriology, virology, nuclear medicine, food industry and paint industry, as well as of infusion flasks, gas bottles and equipment in pharmacies.

Characteristics: neodisher LaboClean FT has a very good cleaning action and is free from detergents including surface active substances. For the removal of nutrient mediums, blood, proteins, radioactive contaminants and residues of tissue cultures. neodisher LaboClean FT has proved to be particularly effective in the residue-free cleaning of tissue-culture flasks, as well as in the area of microbiology and nuclear medicine. Oxidative action from active chlorine.
Laboratory utensils of glass, ceramics, quartz and plastics can be cleaned with neodisher LaboClean FT solutions without affecting the material. Not suitable for non-ferrous metals and light alloys. The use of water with a maximum of 5 °d (1 mmol/l) total hardness is recommended for the cleaning stage.

Application and dosage: 3 - 4 ml/l using suitable dosing devices.

Notes:

For the cleaning of glass culture dishes and infusion bottles:

Cleaning: 3 - 4 ml/l neodisher LaboClean FT at 80 - 85 °C

Neutralisation: 2 - 3 ml/l neodisher Z or neodisher N

For removing food residues:

Pre-rinse: If necessary, with the addition of 0,1 ml/L neodisher EM.

Cleaning: 2 - 3 ml/l neodisher LaboClean FT, if necessary, add 0.1 ml/l neodisher EM

Neutralisation: 1 - 2 ml/l neodisher N

For removing vegetable and fruit residues:

Pre-rinse: 2 ml/l neodisher LaboClean FT cold

Cleaning: 3 - 4 ml/l neodisher LaboClean FT at 60 - 85 °C,

in the event of foaming, add neodisher Defoamer S

Neutralisation: 1 - 2 ml/l neodisher N

For removing textile dyes:

Pre-rinse in the case of alkaline dyestuffs: 1 g/l neodisher Super

Pre-rinse in the case of other dyestuffs: 2 ml/l neodisher LaboClean FT

Cleaning: 3 - 4 ml/l neodisher LaboClean FT at ca. 70 °C

Neutralisation: 1 - 2 ml/l neodisher N

Subsequently the neodisher LaboClean FT- solution has to be rinsed out completely by sufficient rinsing.

Do not mix with other products.

Rinse out dosing system including hose with water before changing product.

Only for professional use.

Technical data: Specific gravity (20 °C): 1.3 g/cm³
pH-range (determined in deionised water, 20 °C) 3 - 4 ml/l: 11.8 - 12.2
Viscosity (concentrate, 20 °C): < 50 mPas
Titration factor: 0.93 (in accordance with neodisher titration method)

Ingredients: Ingredients according to Regulation (EC) No 648/2004 on detergents:
< 5 % chlorine based bleaching agent
5 - 15 % phosphates



Chemische Fabrik Dr. Weigert GmbH & Co. KG
Mühlenhagen 85, D - 20539 Hamburg

Telefon: (040) 789 60 - 0
Telefax: (040) 789 60 - 120

E-Mail: info@drweigert.de
Internet: www.drweigert.de

Die Angaben dieses Merkblattes beruhen auf unseren derzeitigen Kenntnissen und Erfahrungen. Sie befreien den Verwender nicht von eigenen Prüfungen und Versuchen. Eine rechtlich verbindliche Zusicherung bestimmter Eigenschaften kann hieraus nicht abgeleitet werden.


With the above information, which is appropriate to our current knowledge we describe our product regarding possible safety necessities, but we do not involve any quality description or promise certain properties.

neodisher[®] LaboClean FT

Liquid, alkaline cleaning agent
for use in special washing machines



Storage information:

Product is very sensitive to heat and sunlight. Store in a cool place.
Sensitive to frost below - 20 °C.
Expiry date: refer the stamp mark on the label after the symbol 

Hazard warnings and safety advice:

Classification and labelling according to the Dangerous Preparation Directive 1999/45/EC for the concentrate as delivered.

Hazard symbol C - corrosive

contains: Potassium hydroxide, sodium hypochlorite (1 - 5 % active chlorine)

Risk phrases:

- R 31 - Contact with acids liberates toxic gas.
R 35 - Causes severe burns.

Safety phrases:

- S 23 - Do not breath aerosol.
S 26 - In case of contact with eyes, rinse immediately with plenty of water and seek medical advice.
S 27 - Take off immediately all contaminated clothing.
S 28 - After contact with skin, wash immediately with plenty of water.
S 36/37/39 - Wear suitable protective clothing, gloves and eye/face protection.
S 45 - In case of accident or if you feel unwell, seek medical advice immediately (show the label where possible).

If applied according to the instructions for use the product is safe in the sense of the appropriate guidelines upon medical devices.

Dispose only when container is empty and closed. For disposal of product residues, refer to Material Safety Data Sheet.

For further safety information see EC-Safety data sheets. These are available e.g. under www.drweigert.de under the headline "Service".

MB 4100 GB/2-1
01/07



Chemische Fabrik Dr. Weigert GmbH & Co. KG
Mühlenhagen 85, D – 20539 Hamburg

Telefon: (040) 789 60 - 0
Telefax: (040) 789 60 - 120

E-Mail: info@drweigert.de
Internet: www.drweigert.de

Die Angaben dieses Merkblattes beruhen auf unseren derzeitigen Kenntnissen und Erfahrungen. Sie befreien den Verwender nicht von eigenen Prüfungen und Versuchen. Eine rechtlich verbindliche Zusicherung bestimmter Eigenschaften kann hieraus nicht abgeleitet werden.

With the above information, which is appropriate to our current knowledge we describe our product regarding possible safety necessities, but we do not involve any quality description or promise certain properties.